

Patient: NEFTALI ARODI SOLIS ESPINOSA 58A (male, #2508202602)
Referring Physician: IMSS
Study: T.C ABDOMEN SYC (#7363)
Series: Dose Record (#997)
Manufacturer: GE Medical Systems (Optima CT540, #*)
Completion Flag: COMPLETE
Verification Flag: UNVERIFIED
Content Date/Time: 2026-08-25 09:48:51

X-Ray Radiation Dose Report

(observed: 2026-08-25 10:02:33)

Concept Modifier: Procedure reported [[Annex 1](#)]
Observation Context: Observer Type = Device (121007, DCM)
Observation Context: Device Observer UID = 1.2.840.113619.6.278
Observation Context: Device Observer Name = "MWICT540"
Observation Context: Device Observer Manufacturer = "GE Medical Systems"
Observation Context: Device Observer Model Name = "Optima CT540"
Observation Context: Device Observer Serial Number = "*"
Observation Context: Device Role in Procedure = Irradiating Device (113859, DCM)
Observation Context: Device Role in Procedure = Recording (121097, DCM)
Observation Context: Start of X-Ray Irradiation = 2026-08-25 09:48:51
Observation Context: End of X-Ray Irradiation = 2026-08-25 10:02:33
Observation Context: Scope of Accumulation [[Annex 2](#)]

CT Accumulated Dose Data

7 {events} 2661.49 mGy.cm

CT Acquisition

6.3 ABDOMEN DUAL Abdomen Constant Angle Acquisition
1.2.840.113619.2.278.3.2831181000.175.1787659015.301

CT Acquisition Parameters

5.6 s 568.36 mm 64.36 mm -504.0 mm 1.25 mm 5.0 mm 1 {X-Ray sources}

CT X-Ray Source Parameters

1 120.0 kV 10.0 mA 10.0 mA

CT Dose

0.07 mGy IEC Body Dosimetry Phantom 3.85 mGy.cm

Dose Check Alert Details

No Yes 1000.0 mGy

Dose Check Notification Details

No No

CT Acquisition

6.3 ABDOMEN DUAL Abdomen Constant Angle Acquisition
1.2.840.113619.2.278.3.2831181000.175.1787659015.303

CT Acquisition Parameters

5.6 s 568.36 mm 64.36 mm -504.0 mm 1.25 mm 5.0 mm 1 {X-Ray sources}

CT X-Ray Source Parameters

1 120.0 kV 10.0 mA 10.0 mA

CT Dose

0.07 mGy IEC Body Dosimetry Phantom 3.85 mGy.cm

Dose Check Alert Details

No Yes 1000.0 mGy

Dose Check Notification Details

No No

CT Acquisition

6.3 ABDOMEN DUAL Abdomen Spiral Acquisition
1.2.840.113619.2.278.3.2831181000.175.1787659015.306

CT Acquisition Parameters

11.73 s 537.61 mm 509.0 mm 558.34 mm 21.0 mm -488.0 mm 35.52 mm -502.08 mm
1.2.840.113619.2.278.3.2831181000.175.1787659015.299.16810.5 1.25 mm 20.0 mm 1.38 {ratio} 1 {X-Ray sources}

CT X-Ray Source Parameters

1 120.0 kV 440.0 mA 314.8 mA 0.6 s

CT Dose

12.16 mGy IEC Body Dosimetry Phantom 653.9 mGy.cm

Dose Check Alert Details

No Yes 1000.0 mGy

Dose Check Notification Details

No No

CT Acquisition

6.3 ABDOMEN DUAL Abdomen Stationary Acquisition
1.2.840.113619.2.278.3.2831181000.175.1787659015.821

CT Acquisition Parameters

3.6 s 10.0 mm 10.0 mm -13.49 mm -23.49 mm -13.49 mm -23.49 mm
1.2.840.113619.2.278.3.2831181000.175.1787659015.299.16810.5 2.5 mm 10.0 mm 1 {X-Ray sources}

CT X-Ray Source Parameters

1 120.0 kV 100.0 mA 100.0 mA 0.6 s

CT Dose

35.7 mGy IEC Body Dosimetry Phantom 35.7 mGy.cm

Dose Check Alert Details

No Yes 1000.0 mGy

Dose Check Notification Details

No No

CT Acquisition

6.3 ABDOMEN DUAL Abdomen Spiral Acquisition
1.2.840.113619.2.278.3.2831181000.175.1787659015.479

CT Acquisition Parameters

11.73 s 537.61 mm 509.0 mm 558.34 mm 21.0 mm -488.0 mm 35.52 mm -502.08 mm
1.2.840.113619.2.278.3.2831181000.175.1787659015.299.16810.5 1.25 mm 20.0 mm 1.38 {ratio} 1 {X-Ray sources}

CT X-Ray Source Parameters

1 120.0 kV 440.0 mA 314.7 mA 0.6 s

CT Dose

12.16 mGy IEC Body Dosimetry Phantom 653.69 mGy.cm

Dose Check Alert Details

No Yes 1000.0 mGy

Dose Check Notification Details

No No

CT Acquisition

6.3 ABDOMEN DUAL Abdomen Spiral Acquisition
1.2.840.113619.2.278.3.2831181000.175.1787659015.649

CT Acquisition Parameters

11.73 s 537.61 mm 509.0 mm 558.34 mm 21.0 mm -488.0 mm 35.09 mm -502.51 mm
1.2.840.113619.2.278.3.2831181000.175.1787659015.299.16810.5 1.25 mm 20.0 mm 1.38 {ratio} 1 {X-Ray sources}

CT X-Ray Source Parameters

1 120.0 kV 440.0 mA 316.1 mA 0.6 s

CT Dose

12.21 mGy IEC Body Dosimetry Phantom 656.6 mGy.cm

Dose Check Alert Details

No Yes 1000.0 mGy

Dose Check Notification Details

No No

CT Acquisition

6.3 ABDOMEN DUAL Abdomen Spiral Acquisition
1.2.840.113619.2.278.3.2831181000.175.1787659015.865

CT Acquisition Parameters

11.73 s 537.61 mm 509.0 mm 558.34 mm 21.0 mm -488.0 mm 35.52 mm -502.08 mm
1.2.840.113619.2.278.3.2831181000.175.1787659015.299.16810.5 1.25 mm 20.0 mm 1.38 {ratio} 1 {X-Ray sources}

CT X-Ray Source Parameters

1 120.0 kV 440.0 mA 314.8 mA 0.6 s

CT Dose

12.16 mGy IEC Body Dosimetry Phantom 653.9 mGy.cm

Dose Check Alert Details

No Yes 1000.0 mGy

Dose Check Notification Details

No No

Source of Dose Information:

Automated Data Collection (113856, DCM)

Annex

[Annex 1](#)

Procedure reported:

Computed Tomography X-Ray (P5-08000, SRT)

Concept Modifier: Has Intent = Diagnostic Intent (R-408C3, SRT)

[Annex 2](#)

Scope of Accumulation:

Study (113014, DCM)

Properties: Study Instance UID = 1.2.840.113619.2.278.3.2831181000.175.1787659015.297

This page was generated from a DICOM Structured Reporting document by [OFFIS DCMTK](#) 3.6.5.