

Patient: JUANA GARCIA DIAZ 46A (female, #2406202608)
Referring Physician: HGT
Study: ANGIOTAC TORAX (#6833)
Series: Dose Record (#997)
Manufacturer: GE Medical Systems (Optima CT540, #*)
Completion Flag: COMPLETE
Verification Flag: UNVERIFIED
Content Date/Time: 2026-06-24 17:07:23

X-Ray Radiation Dose Report

(observed: 2026-06-24 17:12:51)

Concept Modifier: Procedure reported [[Annex 1](#)]
Observation Context: Observer Type = Device (121007, DCM)
Observation Context: Device Observer UID = 1.2.840.113619.6.278
Observation Context: Device Observer Name = "MWICT540"
Observation Context: Device Observer Manufacturer = "GE Medical Systems"
Observation Context: Device Observer Model Name = "Optima CT540"
Observation Context: Device Observer Serial Number = "*"
Observation Context: Device Role in Procedure = Irradiating Device (113859, DCM)
Observation Context: Device Role in Procedure = Recording (121097, DCM)
Observation Context: Start of X-Ray Irradiation = 2026-06-24 17:07:23
Observation Context: End of X-Ray Irradiation = 2026-06-24 17:12:51
Observation Context: Scope of Accumulation [[Annex 2](#)]

CT Accumulated Dose Data

6 {events} 1636.71 mGy.cm

CT Acquisition

5.2 TORAX ANGIO Chest Constant Angle Acquisition
1.2.840.113619.2.278.3.2831181000.973.1782338213.337

CT Acquisition Parameters

4.1 s 418.36 mm 64.36 mm -354.0 mm 1.25 mm 5.0 mm 1 {X-Ray sources}

CT X-Ray Source Parameters

1 120.0 kV 10.0 mA 10.0 mA

CT Dose

0.07 mGy IEC Body Dosimetry Phantom 2.83 mGy.cm

Dose Check Alert Details

No Yes 1000.0 mGy

Dose Check Notification Details

No No

CT Acquisition

5.2 TORAX ANGIO Chest Constant Angle Acquisition
1.2.840.113619.2.278.3.2831181000.973.1782338213.339

CT Acquisition Parameters

4.1 s 418.36 mm 64.36 mm -354.0 mm 1.25 mm 5.0 mm 1 {X-Ray sources}

CT X-Ray Source Parameters

1 120.0 kV 10.0 mA 10.0 mA

CT Dose

0.07 mGy IEC Body Dosimetry Phantom 2.83 mGy.cm

Dose Check Alert Details

No Yes 1000.0 mGy

Dose Check Notification Details

No No

CT Acquisition

5.2 TORAX ANGIO Chest Spiral Acquisition 1.2.840.113619.2.278.3.2831181000.973.1782338213.342

CT Acquisition Parameters

13.62 s 374.65 mm 359.0 mm 387.34 mm 4.82 mm -354.18 mm 12.77 mm -361.87 mm
1.2.840.113619.2.278.3.2831181000.973.1782338213.335.16630.5 0.62 mm 10.0 mm 1.38 {ratio} 1 {X-Ray sources}

CT X-Ray Source Parameters

1 120.0 kV 400.0 mA 350.2 mA 0.5 s

CT Dose

14.09 mGy IEC Body Dosimetry Phantom 528.05 mGy.cm

Dose Check Alert Details

No Yes 1000.0 mGy

Dose Check Notification Details

No No

CT Acquisition

5.2 TORAX ANGIO Chest Stationary Acquisition

1.2.840.113619.2.278.3.2831181000.973.1782338213.708

CT Acquisition Parameters

4.5 s 10.0 mm 10.0 mm -100.68 mm -110.68 mm -100.68 mm -110.68 mm

1.2.840.113619.2.278.3.2831181000.973.1782338213.335.16630.5 2.5 mm 10.0 mm 1 {X-Ray sources}

CT X-Ray Source Parameters

1 120.0 kV 100.0 mA 100.0 mA 0.5 s

CT Dose

44.63 mGy IEC Body Dosimetry Phantom 44.63 mGy.cm

Dose Check Alert Details

No Yes 1000.0 mGy

Dose Check Notification Details

No No

CT Acquisition

5.2 TORAX ANGIO Chest Spiral Acquisition 1.2.840.113619.2.278.3.2831181000.973.1782338213.466

CT Acquisition Parameters

13.62 s 374.65 mm 359.0 mm 387.34 mm 4.82 mm -354.18 mm 12.77 mm -361.87 mm
1.2.840.113619.2.278.3.2831181000.973.1782338213.335.16630.5 0.62 mm 10.0 mm 1.38 {ratio} 1 {X-Ray sources}

CT X-Ray Source Parameters

1 120.0 kV 400.0 mA 350.1 mA 0.5 s

CT Dose

14.09 mGy IEC Body Dosimetry Phantom 527.9 mGy.cm

Dose Check Alert Details

No Yes 1000.0 mGy

Dose Check Notification Details

No No

CT Acquisition

5.2 TORAX ANGIO Chest Spiral Acquisition 1.2.840.113619.2.278.3.2831181000.973.1782338213.586

CT Acquisition Parameters

13.62 s 374.65 mm 359.0 mm 387.34 mm 4.82 mm -354.18 mm 12.51 mm -362.13 mm
1.2.840.113619.2.278.3.2831181000.973.1782338213.335.16630.5 0.62 mm 10.0 mm 1.38 {ratio} 1 {X-Ray sources}

CT X-Ray Source Parameters

1 120.0 kV 400.0 mA 351.8 mA 0.5 s

CT Dose

14.16 mGy IEC Body Dosimetry Phantom 530.46 mGy.cm

Dose Check Alert Details

No Yes 1000.0 mGy

Dose Check Notification Details

No No

Source of Dose Information:

Annex

Annex 1

Procedure reported:

Computed Tomography X-Ray (P5-08000, SRT)

Concept Modifier: Has Intent = Diagnostic Intent (R-408C3, SRT)

Annex 2

Scope of Accumulation:

Study (113014, DCM)

Properties: Study Instance UID = 1.2.840.113619.2.278.3.2831181000.973.1782338213.333

This page was generated from a DICOM Structured Reporting document by [OFFIS DCMTK](#) 3.6.5.