

Patient: ABEL CAMPUZANO SANCHEZ 69A (male, #1307202604)
Referring Physician: A QUIEN CORRESPONDA
Study: T.C ABDOMEN SYC (#6978)
Series: Dose Record (#997)
Manufacturer: GE Medical Systems (Optima CT540, #*)
Completion Flag: COMPLETE
Verification Flag: UNVERIFIED
Content Date/Time: 2026-07-13 12:51:56

X-Ray Radiation Dose Report

(observed: 2026-07-13 13:05:02)

Concept Modifier: Procedure reported [[Annex 1](#)]
Observation Context: Observer Type = Device (121007, DCM)
Observation Context: Device Observer UID = 1.2.840.113619.6.278
Observation Context: Device Observer Name = "MWICT540"
Observation Context: Device Observer Manufacturer = "GE Medical Systems"
Observation Context: Device Observer Model Name = "Optima CT540"
Observation Context: Device Observer Serial Number = "*"
Observation Context: Device Role in Procedure = Irradiating Device (113859, DCM)
Observation Context: Device Role in Procedure = Recording (121097, DCM)
Observation Context: Start of X-Ray Irradiation = 2026-07-13 12:51:56
Observation Context: End of X-Ray Irradiation = 2026-07-13 13:05:02
Observation Context: Scope of Accumulation [[Annex 2](#)]

CT Accumulated Dose Data

7 {events} 2119.28 mGy.cm

CT Acquisition

6.3 ABDOMEN DUAL Abdomen Constant Angle Acquisition
1.2.840.113619.2.278.3.2831181000.892.1783947511.801

CT Acquisition Parameters

5.6 s 568.36 mm 64.36 mm -504.0 mm 1.25 mm 5.0 mm 1 {X-Ray sources}

CT X-Ray Source Parameters

1 120.0 kV 10.0 mA 10.0 mA

CT Dose

0.07 mGy IEC Body Dosimetry Phantom 3.85 mGy.cm

Dose Check Alert Details

No Yes 1000.0 mGy

Dose Check Notification Details

No No

CT Acquisition

6.3 ABDOMEN DUAL Abdomen Constant Angle Acquisition
1.2.840.113619.2.278.3.2831181000.892.1783947511.803

CT Acquisition Parameters

5.6 s 568.36 mm 64.36 mm -504.0 mm 1.25 mm 5.0 mm 1 {X-Ray sources}

CT X-Ray Source Parameters

1 120.0 kV 10.0 mA 10.0 mA

CT Dose

0.07 mGy IEC Body Dosimetry Phantom 3.85 mGy.cm

Dose Check Alert Details

No Yes 1000.0 mGy

Dose Check Notification Details

No No

CT Acquisition

6.3 ABDOMEN DUAL Abdomen Spiral Acquisition
1.2.840.113619.2.278.3.2831181000.892.1783947511.806

CT Acquisition Parameters

11.99 s 549.61 mm 521.0 mm 570.34 mm 60.93 mm -460.07 mm 75.44 mm -474.16 mm
1.2.840.113619.2.278.3.2831181000.892.1783947511.799.15775.5 1.25 mm 20.0 mm 1.38 {ratio} 1 {X-Ray sources}

CT X-Ray Source Parameters

1 120.0 kV 440.0 mA 242.3 mA 0.6 s

CT Dose

9.36 mGy IEC Body Dosimetry Phantom 514.54 mGy.cm

Dose Check Alert Details

No Yes 1000.0 mGy

Dose Check Notification Details

No No

CT Acquisition

6.3 ABDOMEN DUAL Abdomen Stationary Acquisition
1.2.840.113619.2.278.3.2831181000.892.1783947512.336

CT Acquisition Parameters

4.8 s 10.0 mm 10.0 mm 41.75 mm 31.75 mm 41.75 mm 31.75 mm
1.2.840.113619.2.278.3.2831181000.892.1783947511.799.15775.5 2.5 mm 10.0 mm 1 {X-Ray sources}

CT X-Ray Source Parameters

1 120.0 kV 100.0 mA 100.0 mA 0.6 s

CT Dose

47.6 mGy IEC Body Dosimetry Phantom 47.6 mGy.cm

Dose Check Alert Details

No Yes 1000.0 mGy

Dose Check Notification Details

No No

CT Acquisition

6.3 ABDOMEN DUAL Abdomen Spiral Acquisition
1.2.840.113619.2.278.3.2831181000.892.1783947511.983

CT Acquisition Parameters

11.99 s 549.61 mm 521.0 mm 570.34 mm 60.93 mm -460.07 mm 75.44 mm -474.16 mm
1.2.840.113619.2.278.3.2831181000.892.1783947511.799.15775.5 1.25 mm 20.0 mm 1.38 {ratio} 1 {X-Ray sources}

CT X-Ray Source Parameters

1 120.0 kV 440.0 mA 242.3 mA 0.6 s

CT Dose

9.36 mGy IEC Body Dosimetry Phantom 514.54 mGy.cm

Dose Check Alert Details

No Yes 1000.0 mGy

Dose Check Notification Details

No No

CT Acquisition

6.3 ABDOMEN DUAL Abdomen Spiral Acquisition
1.2.840.113619.2.278.3.2831181000.892.1783947512.157

CT Acquisition Parameters

12.19 s 558.61 mm 530.0 mm 579.34 mm 69.93 mm -460.07 mm 84.01 mm -474.59 mm
1.2.840.113619.2.278.3.2831181000.892.1783947511.799.15775.5 1.25 mm 20.0 mm 1.38 {ratio} 1 {X-Ray sources}

CT X-Ray Source Parameters

1 120.0 kV 440.0 mA 241.1 mA 0.6 s

CT Dose

9.32 mGy IEC Body Dosimetry Phantom 520.37 mGy.cm

Dose Check Alert Details

No Yes 1000.0 mGy

Dose Check Notification Details

No No

CT Acquisition

6.3 ABDOMEN DUAL Abdomen Spiral Acquisition
1.2.840.113619.2.278.3.2831181000.892.1783947512.380

CT Acquisition Parameters

11.99 s 549.61 mm 521.0 mm 570.34 mm 60.93 mm -460.07 mm 75.44 mm -474.16 mm
1.2.840.113619.2.278.3.2831181000.892.1783947511.799.15775.5 1.25 mm 20.0 mm 1.38 {ratio} 1 {X-Ray sources}

CT X-Ray Source Parameters

1 120.0 kV 440.0 mA 242.3 mA 0.6 s

CT Dose

9.36 mGy IEC Body Dosimetry Phantom 514.54 mGy.cm

Dose Check Alert Details

No Yes 1000.0 mGy

Dose Check Notification Details

No No

Source of Dose Information:

Automated Data Collection (113856, DCM)

Annex

[Annex 1](#)

Procedure reported:

Computed Tomography X-Ray (P5-08000, SRT)

Concept Modifier: Has Intent = Diagnostic Intent (R-408C3, SRT)

[Annex 2](#)

Scope of Accumulation:

Study (113014, DCM)

Properties: Study Instance UID = 1.2.840.113619.2.278.3.2831181000.892.1783947511.797

This page was generated from a DICOM Structured Reporting document by [OFFIS DCMTK](#) 3.6.5.